



Government of Pakistan
Pakistan Institute of Medical Sciences
Islamabad.



No.F.1-22/2020-21/MRI/Purchase/PIMS

Islamabad the:- 31 December, 2020.

Incharge HMIS Department,
PIMS, Islamabad.

Subject:- Uploading the Technical Evaluation report on PIMS Website for Purchase of equipment included in project titled Procurement of MRI Equipment for Radiology Department PIMS.

Please find enclosed herewith Technical Evaluation report for the Procurement of Equipment regarding the above cited subject. The detail is as under:-

- 1) 1.5 Tesla Supercon Wide Bore MRI.
- 2) 128 Slice CT Scan.
- 3) DEXA Scan.

Please upload on PIMS website (Hard copy and soft copy is enclosed here with).

Dr. Minhaj-us-Siraj
Joint Executive Director

Copy to:-

- > PS to Executive Director
- > APS to JED
- > Director Purchase



Government of Pakistan
Pakistan Institute of Medical Sciences
Islamabad



TECHNICAL EVALUATION REPORT FOR PURCHASE OF 1.5 TESLA SUPERCON WIDE BORE MRI, LATEST TECHNOLOGY WITH WORKSTATIONS:-

Sr.#	Description	Shirazi Trading Co.	Siemens Health Care	Motora Digionics
1	Make	GE Healthcare	SIEMENS	Phillips
2	Model	Signa Artist	Magnetom SOLA	Ingenia 1.5T Evolution
3	Country of origin: USA, EU and Japan	USA	Germany	Netherland
4		Confirmation letter of 64 or more simultaneous independent RF receiver channels that can be used on single scan is attached. M/s Shirazi Trading Company (Pvt.) Ltd. is technically compliant and Accepted	Confirmation letter of 64 or more simultaneous independent RF receiver channels that can be used on single scan is attached. M/s Siemens Health Care (Pvt.) Ltd. is technically compliant and Accepted	<ul style="list-style-type: none">• Quoted Model not mentioned on attached FDA and CE• Sole agent certificate not attested by embassy/not linked with Phillips Netherlands.• 44mT/m available but not confirmed at 200 slew rate.• 64 or more simultaneous independent RF receiver channels that can be used on single scan is not offered.• One IV pole integrated offered.• Channels offered but elements not traceable (coils detail mentioned in offer but not traceable in data sheet)• Spectroscopy software is not offered by firm Brain & Liver.• M/s Motora Digionics (Pvt.) Ltd. is technically non-compliant and REJECTED on the above mentioned reasons.



Government of Pakistan
Pakistan Institute of Medical Sciences
Islamabad



**TECHNICAL EVALUATION REPORT FOR PURCHASE OF 128 SLICE CT SCAN LATEST TECHNOLOGY
TRUE 64 DETECTOR BASED SYSTEM WITH WORKSTATION**

Sr.#	Description	Medequips Pvt. Ltd.	Shirazi Trading Co.	Motora Digionics	Siemens Health Care
1	Make	Canon Medical Corp, Japan	GE Healthcare, Japan	Phillips Health Care (Zozuho) Co, Ltd, China	Siemens Healthinier
2	Model	Aquilion Prime SP	Revolution EVO	Incisive CT	Somatom Go Top
3	Country of origin: USA, EU and Japan	Japan	Japan	China	Germany
4		M/s Medequips (Pvt.) Ltd. Is technically compliant	Shirazi Trading Co. Technically Non-complained as dual Certificate not attached	Major deficiency of M/s Matora Digionics Pvt. Ltd. a. System is made of China whereas the tender requirement is USA/Europe and Japan only. This bid is not eligible for further evaluation	<ul style="list-style-type: none">M/s Siemens Health Care (Pvt.) Ltd. is technically compliant.



Government of Pakistan
Pakistan Institute of Medical Sciences
Islamabad



TECHNICAL EVALUATION REPORT FOR PURCHASE OF
DEXA SCAN ROTATING C-ARM

SR.#	DESCRIPTION	M/S SHAHCO MEDICAL (PVT.) LTD	M/S SHERAZI TRADING
1	Make	Hologic	GE, Healthcare
2	Model	Horizon A (Latest Computerized)	Lunar Prodigy Prime China Made (Parts from France, USA, Japan)
3	Country of origin: USA, EU and Japan	USA	Model name not mentioned on attached FDA & CE Certificates
4	Certification: FDA, JIS & CE (at least Two)	FDA and CE Certificate attached	<ul style="list-style-type: none">• Full body rotating C-Arm (fixed table and movable C-arm). Only left to right movement motorized (rotation not available)• Table (flex table with moving C-arm)• Whole body 15 sec to 200 sec or less (N/A) 360 sec total body scan time. M/s Shirazi Trading Company (Pvt.) Ltd. is technically non-compliant and REJECTED on the above mentioned reasons.
5		M/s Shahco Medical (Pvt.) Ltd. is technically compliant.	