



Corrigendum Notice No.02/2026-27/PIMS

Reference to this Institute **Tender Notice No. 02/2026-2027/PIMS**, published in the newspapers (Nations Islamabad and Daily Nawa-e-Waqat Islamabad etc) on **26-05-2026**, the following amendment/ change has been made as under:-

Purchase of Bi-Plane Angiography Machine mentioned in tender Sr. # 06, which was originally scheduled to be opened on 18-06-2026 , has been extended. <u>The tender will now be opened on 30-06-2026 at 11:30 AM in the Committee Room, Purchase Department, PIMS Islamabad.</u>	
<u>Specifications:-</u> mentioned in Sr.1.2 <u>Pixel size:-</u>	<u>May be read as:-</u> <u>≤155µm instead of ≤150µm.</u>
<u>Specifications:-</u> mentioned in Sr.2.4 <u>Software packages:-</u>	<u>The below mentioned points may also be include in software packages with already advertised specifications:-</u> • 3D reconstruction; possibility of visualizing both vessels and other material such as coils, stunts with destructive colors. • 3D road map; overlay of 3D data set on live fluoro with automatic compensation of C-Arm and table movement. • CT like imaging coverage of at least 150 to 200 degrees so that important information is not missed display/ overlay between 3D model and CT like image for visualization and comparison. Metal Artifact reduction. • Fusion with important DICOM Volumes:- Fusion with MR, CT, Imported DICOM images. Fused image can be overlaid / display on live Fluoro. • 3D parenchymal color blood flow imaging (Neuro Perfusion application); color coded imaging of the parenchymal blood volume in the brain during an intervention. • 3D volumetric analysis of cerebral aneurysm automatic calculation of dome height, ostium diameter, and aneurysm volume.

The Corrigendum Notice is also available at **PPRA website** www.ppra.org.pk as well as at PIMS Website www.pims.gov.pk.

Project Director
Ph. 051-9260490